

ICAR-INDIAN AGRICULTURAL RESEARCH INSTITUTE REGIONAL
STATION, KATRAIN, KULLU, HP, INDIA-175 129

F. No. 31/620.....

Dated: 06-03-2019

Subject: Online quotation for the purchase/supply of -80 °C Refrigerator/Deep Freezer at ICAR-Indian Agricultural Research Institute Regional Station, Katrain, Kullu, HP, India-175 129

Online Quotation are invited from reputed manufacturer/Dealer/Supplier in two bid system (Technical bid & Financial bid) for the purchase/supply of -80 °C Refrigerator/Deep Freezer on behalf of Director, ICAR-Indian Agricultural Research Institute, New Delhi. The last date for online submission of Bid is 27.03.2019 at 1:00 pm. The bid will be opened on dated 28.03.2019 at 3:00 pm.

Quotation shall satisfy the following conditions:

1. The rates quoted shall be valid for a minimum period of 180 days from the last date fixed for the online submission of bid.
2. The rates quoted shall be for free delivery and installation at ICAR-Indian Agricultural Research Institute Regional Station, Katrain, Kullu, HP, India-175 129. Costs of imported item/equipment manufactured in abroad, should be quoted in the currency of manufacturing country on FOB basis and showing price up to CIF. Quotation in Indian Rupees will be accepted if the good/equipment is manufactured in India, otherwise the bid will be liable to rejection.
3. Custom Duty Exemption Certificate (CDEC) will be issued only when the bid is quoted in foreign currency in case of foreign made items only and meant for IARI.
4. In case the bid is offered in Indian Rupees on FOR value basis than CDEC/DSIR certificate will not be given by the institute at all. However, VAT as applicable will be paid by the Institute as per rule.
5. The successful bidder will have to provide the original proforma invoice from the foreign principal (duly signed in ink), in case of imported item/equipment, within 15 days of the date of receipt of purchase order, otherwise the purchase order will automatically stand cancelled without any further communication.
6. Full specification of the article quoted for shall be given in the quotation along with literature/pamphlet/drawing etc. User list of the quoted document may also be additionally provided for reference.
7. If taxes, duties or any other charges over and above the rates quoted liveable, actual percentage of such taxes/duties/other charges should be clearly indicated.
8. If the bid is submitted by the authorized dealer on behalf of the Manufacturer/OEM, than valid authorization certificate of the Manufacturer/OEM must be attached with the bid, otherwise bid will not be considered.
9. As per CVC's guide lines No. 12-02-6 CTE-SPJ (1)-2/161730 (Circular No. 3/1/12 dated 13/01/12, in a tender, either the Indian Agent on behalf of the Principal/OEM or Principal/OEM itself can bid but both cannot bid simultaneously for the same item/product in the same tender. Further if an agent submits bid on behalf of the Principal/OEM the same agent shall not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product.
10. Supply is to be made by any convenient mode of transport at supplier's risk unless otherwise specified.
11. In case the material is supplied through an authorized agent, manufacturer shall be responsible for the successful installation of the equipment. The manufacturer shall be liable to any penal action for the shortcomings of the authorized agent. Any payment that is to be made separately to the agent if any shall be clearly mentioned in the quotation.

12. The bid should be accompanied with earnest money of Rs. 20,000/- (Rupees Twenty Thousand only) in the form of D.D. /Pay Order/Fixed Deposit Receipt/Bank Guarantee from commercial bank drawn in favour of Director, IARI. The original EMD should be submitted to the tender inviting authority on or before last date and time of online bid submission. The detail of DD/Pay order/FDR/Bank Guarantee physically sent should tally with the details available in the scanned copy and data entered during the bid submission time otherwise the uploaded bid will be rejected.
13. No interest on security deposit and earnest money deposit shall be paid by the institute to tenderer.
14. The EMD shall be refunded to the unsuccessful bidders after finalization of the quotation. In case of successful bidder, it can be adjusted towards security deposit which is 5-10% of the order value in the form of D.D./Pay order/Fixed Deposit Receipt/Bank Guarantee from commercial bank drawn in favour of Director, IARI and shall remain valid for 60 days beyond the date of completion of all contractual obligation of supplier including warranty obligation for the equipment/goods costing Rs. 1.00 Lakh (Rs. One Lakh) or more. Here would however, be no performance security deposit for equipment/goods costing less than Rs. 1.00 Lakh (Rs. One Lakh).
15. One-year warranty has to be invariable provide by the firm. In case, there is any variation in the warranty period given in the specification of the item/equipment, the warranty period shall be the final as given in the specification.
16. Payment will be made by mode of e-payment to the supplier/firm after satisfactory completion of work and receipt of pre-receipt bills triplicate. In case of foreign currency payment shall generally be made through LC directly to the manufacturer.
17. The Director IARI, New Delhi reserves the right to accept or reject any or all the quotations either in full or in parts without assigning any reason.
18. Quotations not complying with the above conditions are liable to be rejected.
19. Original EMD shall be send/deposited at administrative office at **Indian Agricultural Research Institute, Regional Station, Katrain-175129 [Himachal Pradesh]**.
20. TIN/VAT/PAN/CST /ISO Nos. may be given on the Quotation.
21. AMC/CMC charges shall not be included for the evaluation of the financial/commercial bids.
22. Financial Bid will be opened of those firms who qualify in Technical bid.
23. There will be no exemption for paying earnest money. However, Firms registered with the Central Purchase Organisation/NSIC only are exempted submitting bids without EMD. (Scanned copy of registration must be attached therewith without which it will not be considered)
24. In case, any dispute arising out of this contract shall be subject to the jurisdiction of Indian laws & court, at New Delhi. Sole arbitrator is appointed by the Secretary, ICAR, New Delhi. His decision will be final and binding to both parties (supplier and Purchaser).
25. The successful bidder has to execute the indented material as mentioned in the supply order placed with them within the stipulated period as given in the supply order placed by the unit/Division of this Institute failing which 2% per week and maximum of 10% deduction as liquidated damage charges will be made from the bill in case the material is not supplied within the given stipulated period.

The scanned copy of following documents must be furnished with the technical bid failing which tender will not be accepted at all.

1. Scanned copy of TIN/VAT/LST/CST.
2. Scanned copy of Pan Card.
3. Scanned copy of current authorization letter issued from manufacturer/principal Firm in original.
4. Scanned copy of Earnest Money Deposit (EMD) or scanned copy of registration from Central Purchase Organization/NSIC.
5. Scanned copy of literature/pamphlet/drawing.
6. Scanned copy of user list.
7. CE, European/ USA certified/RoHS, & ISO certificate.
8. Scanned copy of compliance statement in accordance with the tender technical specification.
9. Scanned copy of the tender documents duly signed by the concerned firm accepting the same.

Description of the item	Qty. Required	Cost not to be exceeded
As per Annexure-I Attached	one	Rs. 4.00 lacks

Copy uploaded on IARI website: www.iari.res.in & CPP Portal (www.eprocure.gov.in)



Asstt. Admn. Officer
ICAR-IARI Regional Station,
Katrain, Kullu, HP, India-175129

Specification for -80 °C Refrigerator/Deep Freezer

1. It should be vertical in shape with a capacity of 185 litres or more.
2. Operating temperature should range between -45 °C to -86 °C.
3. It should have at least 3 shelves with minimum 2 internal doors.
4. Microprocessor based temperature controller with PT-100 sensor.
5. It should have at least 4-inch LCD or touch screen display.
6. Audio visual alarm in case of power failure and door opening.
7. Noise level should be less than 65 db.
8. Battery backup of minimum 8 hrs for temperature control.
9. Power supply: 220-240 volts, 50 Hz in single phase.
10. It should be equipped with 6.0 KVA voltage stabilizer.
11. Warranty period: 2 year from the date of installation including all spares, accessories & labour cost.
12. Should be CE, European/ USA certified/RoHS, & ISO certification should be enclosed with Technical bid.
13. Quoted Model should have at least 10 Nos. of installations in different Indian Institutes.
14. Quotations should be quoted on original brochure and no computer-generated brochure will be accepted.

Note: All the claims and specification about the instrument should be supported by a printed, online and technical literature