
 <p>भा.कृ.अ.प. ICAR</p>	<p>पादप कार्यिकी संभाग DIVISION OF PLANT PHYSIOLOGY भा.कृ.अ.प.—भारतीय कृषि अनुसंधान संस्थान ICAR-Indian Agricultural Research Institute नई दिल्ली—110012 (भारत) NEW DELHI - 110 012 (INDIA)</p>	
--	---	---

फाईल सं०: 51-6 / TCR / GET / 12-197-B


दिनांक: 13 / 02 / 2025

GeM के माध्यम से ई-खरीद हेतु सूचना
NOTICE FOR E-PROCUREMENT THROUGH GeM

Online GeM e- bid are invited from reputed Manufacturere/Supplier/Authorized dealer in two bid system (Technical and Financial) for purchase of scientific equipment **Tissue Culture Racks** on behalf of Director, ICAR-Indian Agricultural Research Institute, New Delhi. Please visit www.iari.res.in for details Rules and Regulation and log in www.gem.gov.in for online e-bidding.

जैम बिड का विवरण निम्नलिखित है / Details of Gem Bid are as follows:

जैम बिड संख्या / GeM Bid No.	GEM/2025/B/5924607
बिड जमा करवाने की तिथि एवं समय / Bid Submission start date and time	07-02-2025
बिड बंद होने की तिथि एवं समय / Last Date & Time for submission of bid	28-02-2025 17.00
बिड खुलने की तिथि एवं समय / Date & Time for opening of Technical Bid	28-02-2025 17.30


प्रशासनिक अधिकारी
Admn.Officer

REQUISITION OF THE PURCHASE OF STORES

Tissue Culture Racks

Technical Specifications:

1. The rack should be made of mild steel powder coated or stainless steel 304.
2. The number of shelves should be 4 or more.
3. Each rack of size 48"x12" (depth) or better.
4. Distance between each rack is 15" or more.
5. Shelves should be provided with Fluorescent / LED tubes ^{Amul} ~~or more~~
6. Should have UV germicidal light
7. 0-24 Hours Digital Timers should be provided in tissue culture rack.
8. The rack should have castor wheels for easy movement
9. Input Voltage- 200-240 V AC 50 Hz single Phase.
10. Should have control panels for turning the lights on and off
11. Switches should be provided to operate each shelf lights.

Special conditions:

12. Bidder should have service centre in Delhi.
13. The inspection certificate duly signed by the attending officer must be submitted as tender document.
14. The OEM should attach EPR registration certificate certified with CPCB for electrical & electronic parts.